



Declaration of Conformity

PRODUCT IDENTIFICATION	
Product name	Model Number
MediCompass Connect (MetrikLink)	ML1000 ML1000USB ML1100 (MetrikLink Model Numbers)

MANUFACTURER		
Name of company	Address	Representative
iMetrikus, Inc.	640 W. California Ave., Suite 200 Sunnyvale, CA 94086 USA	Michael Acosta Vice President Quality Systems & Regulatory Affairs

AUTHORIZED REPRESENTATIVE		
Name of company	Address	Telephone/email
Emergo Europe	Molenstraat 15 2513 BH The Hague, Netherlands	+31.70.345.8570 - phone +31.70.346.7299 - fax service@emergogroup.com

CONFORMITY ASSESSMENT		
Device classification	Route to compliance	Standards applied
Class 1 Rule 12	Annex VII of MDD 93/42/EEC Council Directive	IEC 60601-1; IEC 60950-1; EN 55011; FCC Parts 68 & 15; TBR21.

iMetrikus declares that the above mentioned products meet the provision of the Council Directive 93/42/EEC for Medical Devices and Directive 93/42/EEC as transposed in the national laws of the Member States.

COMPANY REPRESENTATIVE: Michael Acosta

TITLE: Vice President QA/RA

SIGNATURE:

DATE: 03/09/2009